

## **Endoscope Storage Cabinets**





ScopeVAULT™

## SAFETY AND COMPLIANCE

**Scope**VAULT

ONE

- Endoscopes hang freely and don't touch each other, the back wall, or the floor
- Bright LED lights illuminate inventory
- 24/7 HEPA maximum airflow positive pressure
- No floor level vents to draw in dust and dirt.
- Removable drip tray, easy to clean

- Earthquake compliant
- Free standing
- Space saving technology small footprint
- Joint Commission inspection friendly
- Meets or exceeds all current standards
  ANSI®, AAMI® ST91:2015, AORN®, and SGNA®
  compliant
- Modular design is upgradeable to meet future guidelines



Clinical Choice, LLC 204 Muirs Chapel Road Suite 301Q Greensboro, NC 27410

Phone: 877.572.9985 Fax: 877.572.9986 www.clinicalchoice.com





## Got Compliance? We've Got You Covered.

Guidelines	ScopeVAULT™	SGNA®	AORN <sup>®</sup>	AAMI®
Clean, well-ventilated and dust free	$\otimes$		<b>⊘</b>	<b>⊘</b>
Endoscopes should not touch each other (hang freely)	$\otimes$	<b>⊘</b>	<b>⊘</b>	<b>⊘</b>
Endoscopes should not touch the bottom of the cabinet	$\otimes$	-	<b>⊘</b>	<b>⊘</b>
Endoscopes should not be coiled*	$\otimes$	<b>⊘</b>	<b>⊘</b>	<b>⊘</b>
Cabinets should have venting that allows air circulation around the endoscopes	$\otimes$	<b>⊘</b>	<b>⊘</b>	<b>⊘</b>
Sufficient space between and around scopes	$\otimes$	<b>⊘</b>	<b>⊘</b>	<b>⊘</b>
Cabinets should be made of a material that can be disinfected	$\otimes$	<b>⊘</b>	<b>⊘</b>	<b>⊘</b>
Endoscopes should hang in a vertical position*	$\otimes$		<b>⊘</b>	<b>⊘</b>
Area temperature and humidity should be monitored	<b>⊘</b>	-	-	<b>⊘</b>
Flexible and semi-rigid endoscopes should be stored in a closed cabinet	$\otimes$	-	<b>⊘</b>	-

<sup>\*</sup>Per OEM IFU







Maximum Airflow



**Small Footprint** 



Modular Design



Infection Control

ScopeVAULT™ Cabinet Specifications		
SIZE	W 29", D 23", H 94"	
WEIGHT	250+/- pounds	
VOLTAGE	120V 20A Circuit	